

Visy Pulp and Paper Tumut CEMS - Correction Log

14/08/2023

Reporting Period: 1/07/2023 - 1/08/2023

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere


Opacity		Period: 6 Minutes	Limit: 20.00 %			
Start Time	End Time	Cause	Operational State	Explanation	Max Reading	
01/07/23 15:06	01/07/23 15:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.77	
02/07/23 15:06	02/07/23 15:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.17	
03/07/23 15:06	03/07/23 15:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.89	
04/07/23 15:06	04/07/23 15:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.54	
05/07/23 15:06	05/07/23 15:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.54	
06/07/23 15:06	06/07/23 15:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.52	
07/07/23 15:06	07/07/23 15:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.03	
08/07/23 15:06	08/07/23 15:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.85	
13/07/23 15:00	13/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.44	
14/07/23 15:00	14/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.85	
15/07/23 15:00	15/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.91	
16/07/23 15:00	16/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.97	
17/07/23 15:00	17/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.58	
18/07/23 15:00	18/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.32	
19/07/23 15:00	19/07/23 15:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	31.04	
20/07/23 15:00	20/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.07	
21/07/23 15:00	21/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.69	
22/07/23 15:00	22/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.15	

23/07/23 15:00	23/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.01
24/07/23 15:00	24/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.37
25/07/23 15:00	25/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.79
26/07/23 15:00	26/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.16
27/07/23 15:00	27/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.53
28/07/23 15:00	28/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.46
29/07/23 15:00	29/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.75
30/07/23 15:00	30/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	27.32
31/07/23 15:00	31/07/23 15:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.06

Main Stack 2

Monitoring Location No: 22
Monitoring Type: Continuous
Sample Type: Air
Description: Exit point from Stack 2 to atmosphere

Hydrogen Chloride (HCl) Period: 60 Minutes Limit: 50.00 mg/Nm3					
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
14/07/23 09:00	14/07/23 10:00	Instrument Error	Normal (Steady State)	Friday Maintenance calibration carried out	85.27
28/07/23 10:00	28/07/23 11:00	Instrument Error	Normal (Steady State)	Friday Maintenance calibrations carried out	71.33

Authorised By: 

Uday Bhagwat
Pulp Mill Manager


Johan Stoltz
General Manager