

Main Stack 1


Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %			
Start Time	End Time	Cause	Operational State	Explanation	Max Reading	
01/10/23 15:42	01/10/23 15:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.56	
02/10/23 15:42	02/10/23 15:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.99	
03/10/23 15:42	03/10/23 15:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.76	
04/10/23 15:42	04/10/23 15:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.04	
05/10/23 15:42	05/10/23 15:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.34	
06/10/23 15:42	06/10/23 15:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.25	
07/10/23 15:42	07/10/23 15:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.69	
08/10/23 15:42	08/10/23 15:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	31.60	
08/10/23 17:18	08/10/23 17:36	Instrument Error	Lime Kiln B Scheduled Start-up/Shut-down	Mill shutting down for MSD 2023	27.54	
09/10/23 15:42	09/10/23 15:48	Lime Kiln B	Lime Kiln B Scheduled Start-up/Shut-down	Mill shutting down for MSD 2023	34.18	
10/10/23 15:42	10/10/23 15:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	27.35	
17/10/23 15:36	17/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.54	
18/10/23 15:36	18/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.30	
19/10/23 15:36	19/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.00	
20/10/23 15:36	20/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.83	
21/10/23 15:36	21/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.17	
22/10/23 15:36	22/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.97	
23/10/23 15:36	23/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.83	

24/10/23 15:36	24/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.65
25/10/23 15:36	25/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.05
26/10/23 15:36	26/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.89
27/10/23 15:36	27/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.70
28/10/23 15:36	28/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.30
29/10/23 15:36	29/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.96
30/10/23 15:36	30/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.75
31/10/23 15:36	31/10/23 15:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.51

Authorised By:

Uday Bhagwat
Pulp Mill Manager



Johan Stoltz
General Manager

