



# Visy Pulp and Paper Tumut CEMS - Correction Log

10/01/2024

Reporting Period: 1/12/2023 - 1/01/2024

Environment Protection Licence No: 10232

## Main Stack 1

Monitoring Location No: 1  
 Monitoring Type: Continuous  
 Sample Type: Air  
 Description: Exit point from Stack 1 to atmosphere

Opacity						Period: 6 Minutes	Limit: 20.00 %
Start Time	End Time	Cause	Operational State	Explanation	Max Reading		
04/12/23 15:24	04/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.22		
05/12/23 15:24	05/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.29		
06/12/23 15:24	06/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.07		
07/12/23 15:24	07/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.12		
08/12/23 15:24	08/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.63		
09/12/23 15:24	09/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.95		
10/12/23 15:24	10/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.46		
11/12/23 15:24	11/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.13		
12/12/23 15:24	12/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.67		
13/12/23 15:24	13/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.77		
14/12/23 15:24	14/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.25		
15/12/23 15:24	15/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.15		
16/12/23 15:24	16/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.93		
17/12/23 15:24	17/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.24		
18/12/23 15:24	18/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.84		
19/12/23 15:24	19/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.23		
20/12/23 15:24	20/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.79		
21/12/23 15:24	21/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.30		

22/12/23 15:24	22/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.70
23/12/23 15:24	23/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.64
24/12/23 15:24	24/12/23 15:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.51
29/12/23 15:18	29/12/23 15:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.40
30/12/23 15:18	30/12/23 15:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.75
31/12/23 15:18	31/12/23 15:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.34

Authorised By:

**Uday Bhagwat**  
Pulp Mill Manager



**Johan Stoltz**  
General Manager

