



Visy Pulp and Paper Tumut CEMS - Correction Log

6/06/2024

Reporting Period: 1/05/2024 - 1/06/2024

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity					
Period: 6 Minutes		Limit: 20.00 %			
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
01/05/24 08:00	01/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.10
02/05/24 08:00	02/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.97
03/05/24 08:00	03/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.96
04/05/24 08:00	04/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.25
05/05/24 08:00	05/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.24
06/05/24 08:00	06/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.34
07/05/24 08:00	07/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.94
08/05/24 08:00	08/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.35
09/05/24 08:00	09/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.63
10/05/24 08:00	10/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.87
11/05/24 08:00	11/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.24
12/05/24 08:00	12/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.61
13/05/24 08:00	13/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.56
15/05/24 08:00	15/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.27
16/05/24 07:54	16/05/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	45.66
18/05/24 07:54	18/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.05
19/05/24 07:54	19/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.70
20/05/24 07:54	20/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.98

21/05/24 07:54	21/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.76
22/05/24 07:54	22/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.27
23/05/24 07:54	23/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.16
24/05/24 07:54	24/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.84
25/05/24 07:54	25/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.18
26/05/24 07:54	26/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.78
27/05/24 07:54	27/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.76
28/05/24 07:54	28/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.31
29/05/24 07:54	29/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.69
30/05/24 07:54	30/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.78
31/05/24 07:54	31/05/24 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.00

TRS (as H2S) Period: 60 Minutes Limit: 3.60 mg/Nm3					
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
10/05/24 10:00	10/05/24 11:00	Instrument Error	Normal (Steady State)	Commissioning new TRS installation - New equipment installed and tested: Dilution Probe Heater failure - no response to controller Re-installed old probe, Also Failed - System not sampling from stack.	12.96
28/05/24 14:00	28/05/24 15:00	Instrument Error	Normal (Steady State)	Replacing the TRS unit - planned maintenance - no true exceedance	6.60
29/05/24 09:00	29/05/24 10:00	Instrument Error	Normal (Steady State)	New TRS monitoring equipment installed.	20.67
29/05/24 11:00	29/05/24 13:00	Instrument Error	Normal (Steady State)	New TRS monitoring equipment installed.	8.70

Main Stack 2

Monitoring Location No: 22
Monitoring Type: Continuous
Sample Type: Air
Description: Exit point from Stack 2 to atmosphere

Nitrogen Oxides (as NO2) Period: 60 Minutes Limit: 400.00 mg/Nm3					
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
17/05/24 13:00	17/05/24 14:00	Instrument Error	Normal (Steady State)	Weekly CEMS check by Instrumentation team. Not a true exceedance.	453.41

Opacity					
Period: 6 Minutes		Limit: 20.00 %			
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
09/05/24 08:18	09/05/24 08:30	Instrument Error	Normal (Steady State)	The opacity instrument was inspected by Instrumentation service provider. No real exceedance.	26.39

Authorised By:

Uday Bhagwat
Pulp Mill Manager

Johan Stoltz
General Manager

