



Visy Pulp and Paper Tumut CEMS - Correction Log

15/08/2024

Reporting Period: 1/07/2024 - 1/08/2024 Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity						Period: 6 Minutes	Limit: 20.00 %
Start Time	End Time	Cause	Operational State	Explanation	Max Reading		
05/07/24 07:42	05/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.26		
06/07/24 07:42	06/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.74		
07/07/24 07:42	07/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.96		
08/07/24 07:42	08/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.42		
09/07/24 07:42	09/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.91		
10/07/24 07:42	10/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.95		
11/07/24 07:42	11/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.30		
12/07/24 07:42	12/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.14		
13/07/24 07:42	13/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.41		
14/07/24 07:42	14/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.08		
15/07/24 07:42	15/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.69		
16/07/24 07:42	16/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.25		
17/07/24 07:42	17/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.98		
18/07/24 07:42	18/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	27.96		
20/07/24 07:42	20/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	27.04		
23/07/24 07:42	23/07/24 07:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.58		
28/07/24 07:36	28/07/24 07:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.06		
29/07/24 07:36	29/07/24 07:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.60		

30/07/24 07:36	30/07/24 07:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.63
31/07/24 07:36	31/07/24 07:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.18

Main Stack 2

Monitoring Location No: 22
Monitoring Type: Continuous
Sample Type: Air
Description: Exit point from Stack 2 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
20/07/24 10:18	20/07/24 10:24	Instrument Error	Equipment Issue/Failure	Main stack Opacity Instrument did 1 self-check then failed with the Windows crash, was powered down and then rebooted by E&I restoring the value.	22.48

Authorised By:

Uday Bhagwat
 Pulp Mill Manager



Johan Stoltz
 General Manager

