



# Visy Pulp and Paper Tumut CEMS - Correction Log

13/11/2024

Reporting Period: 1/10/2024 - 1/11/2024

Environment Protection Licence No: 10232

## Main Stack 1

Monitoring Location No: 1

Monitoring Type: Continuous

Sample Type: Air

Description: Exit point from Stack 1 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
01/10/24 07:24	01/10/24 07:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.71
02/10/24 07:24	02/10/24 07:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.15
08/10/24 08:18	08/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.03
09/10/24 08:18	09/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.88
10/10/24 08:18	10/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.03
11/10/24 08:18	11/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.79
12/10/24 08:18	12/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.90
13/10/24 08:18	13/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.31
14/10/24 08:18	14/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.91
15/10/24 08:18	15/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.29
19/10/24 02:42	19/10/24 02:54	Instrument Error	Equipment Issue/Failure	Mill still under MSD2024 No process plants started yet.	34.35
19/10/24 08:18	19/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.36
20/10/24 08:18	20/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.55
21/10/24 08:18	21/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.00
21/10/24 14:00	21/10/24 14:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	34.95
22/10/24 08:18	22/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.60
23/10/24 08:18	23/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	27.83

24/10/24 08:18	24/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.48
25/10/24 08:18	25/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.12
26/10/24 08:18	26/10/24 08:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.16

TRS (as H2S)      Period: 60 Minutes      Limit: 3.60 mg/Nm3					
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
17/10/24 21:00	17/10/24 22:00	Instrument Error	Equipment Issue/Failure	Mill shut down MSD2024	4.65
18/10/24 00:00	18/10/24 03:00	Instrument Error	Equipment Issue/Failure	Mill shut down MSD2024	21.27
18/10/24 05:00	18/10/24 07:00	Instrument Error	Equipment Issue/Failure	MSD2024 all Plants on shut	17.09

## Main Stack 2

**Monitoring Location No:** 22  
**Monitoring Type:** Continuous  
**Sample Type:** Air  
**Description:** Exit point from Stack 2 to atmosphere

Hydrogen Chloride (HCl)      Period: 60 Minutes      Limit: 50.00 mg/Nm3					
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
04/10/24 08:00	04/10/24 09:00	Instrument Error	Equipment Issue/Failure	Weekly Maintenance work done on CEMS	154.87

Opacity      Period: 6 Minutes      Limit: 20.00 %					
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
19/10/24 08:00	19/10/24 08:12	Instrument Error	Equipment Issue/Failure	Boiler offline MSD2024	20.82
19/10/24 10:54	19/10/24 11:00	Instrument Error	Equipment Issue/Failure	Boiler offline - MSD2024	23.14





Authorised By:

**Uday Bhagwat**  
Pulp Mill Manager



---

**Johan Stoltz**  
General Manager



---