



Visy Pulp and Paper Tumut CEMS - Correction Log

8/01/2025

Reporting Period: 1/12/2024 - 1/01/2025

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %			
Start Time	End Time	Cause	Operational State	Explanation	Max Reading	
01/12/24 08:06	01/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.67	
02/12/24 08:06	02/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.22	
03/12/24 08:06	03/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.10	
04/12/24 08:06	04/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.17	
05/12/24 08:06	05/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.80	
06/12/24 08:06	06/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.46	
07/12/24 08:06	07/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.35	
08/12/24 08:06	08/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.87	
09/12/24 08:06	09/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.19	
10/12/24 08:06	10/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.68	
11/12/24 08:06	11/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.84	
12/12/24 08:06	12/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.84	
13/12/24 08:06	13/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.75	
14/12/24 08:06	14/12/24 08:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.54	
20/12/24 08:00	20/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.17	
21/12/24 08:00	21/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.59	
22/12/24 08:00	22/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.95	
23/12/24 08:00	23/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.55	

24/12/24 08:00	24/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.25
25/12/24 08:00	25/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.75
26/12/24 08:00	26/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.14
27/12/24 08:00	27/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.45
28/12/24 08:00	28/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.59
29/12/24 08:00	29/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.85
30/12/24 08:00	30/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.34
31/12/24 08:00	31/12/24 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.57

Power Boiler

Monitoring Location No: 3
Monitoring Type: Continuous
Sample Type: Air
Description: Discharge duct downstream of Power Boiler prior to junction with Stack 1

Carbon Monoxide (CO)		Period: 60 Minutes	Limit: 140.00 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
13/12/24 05:00	13/12/24 08:00	Instrument Error	Normal (Steady State)	Sensor blocking up with dust (prematurely). Removed sensor and cleaned it - working again. Suspect the shielding is eroded away	169.72
13/12/24 09:00	13/12/24 11:00	Instrument Error	Normal (Steady State)	Sensor blocking up with dust (prematurely). Removed sensor and cleaned it - working again. Suspect the shielding is eroded away	215.69
19/12/24 19:00	19/12/24 20:00	Instrument Error	Normal (Steady State)	The CO probe gets dirty and seems to need a weekly clean to be reliable. This is not a true exceedance.	142.07

Main Stack 2

Monitoring Location No: 22
Monitoring Type: Continuous
Sample Type: Air
Description: Exit point from Stack 2 to atmosphere

Nitrogen Oxides (as NO ₂)		Period: 60 Minutes	Limit: 400.00 mg/Nm ³		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
10/12/24 18:00	11/12/24 05:00	Instrument Error	Equipment Issue/Failure	CEMS unit issues, struggling to do calibrations.	813.08
11/12/24 00:00	11/12/24 01:00	Instrument Error	Equipment Issue/Failure	CEMS unit issues, struggling to do calibrations.	770.12
11/12/24 05:00	11/12/24 08:00	Instrument Error	Normal (Steady State)	Maintenance worked on CEMS unit with assistance from Leir Siegler to sort out calibration issues.	781.34

Authorised By:

Uday Bhagwat
Pulp Mill Manager



Johan Stoltz
General Manager



