



Visy Pulp and Paper Tumut CEMS - Correction Log

7/02/2025

Reporting Period:

1/01/2025 - 1/02/2025

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
01/01/25 08:00	01/01/25 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.20
02/01/25 08:00	02/01/25 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.23
03/01/25 08:00	03/01/25 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.66
04/01/25 08:00	04/01/25 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.66
05/01/25 08:00	05/01/25 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.13
06/01/25 08:00	06/01/25 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.72
07/01/25 08:00	07/01/25 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.79
08/01/25 08:00	08/01/25 08:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.65
14/01/25 07:54	14/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.52
15/01/25 07:54	15/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.92
16/01/25 07:54	16/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.13
17/01/25 07:54	17/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.16
18/01/25 07:54	18/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.35
19/01/25 07:54	19/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.17
20/01/25 07:54	20/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.28
21/01/25 07:54	21/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.62
22/01/25 07:54	22/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.97
23/01/25 07:54	23/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.13

24/01/25 07:54	24/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.11
25/01/25 07:54	25/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.37
26/01/25 07:54	26/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.57
27/01/25 07:54	27/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.37
28/01/25 07:54	28/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.26
29/01/25 07:54	29/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.83
30/01/25 07:54	30/01/25 08:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.57

Main Stack 2

Monitoring Location No: 22
Monitoring Type: Continuous
Sample Type: Air
Description: Exit point from Stack 2 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
15/01/25 14:54	15/01/25 15:00	Instrument Error	Equipment Issue/Failure	During rain and electric storm Opacity went up	30.75
23/01/25 12:24	23/01/25 12:30	Instrument Error	Equipment Issue/Failure	Main Stack Opacity was reading much higher than the 2 ducting opacities going into main stack.	27.27

TRS (as H2S)		Period: 60 Minutes	Limit: 3.60 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
30/01/25 14:00	30/01/25 15:00	Instrument Error	Equipment Issue/Failure	Maintenance carried out on CEMS unit by E&I	4.74

Authorised By:

Uday Bhagwat
Pulp Mill Manager



Johan Stoltz
General Manager

