



Visy Pulp and Paper Tumut CEMS - Correction Log

2/02/2022

Reporting Period: 1/01/2022 - 1/02/2022

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
01/01/22 18:24	01/01/22 18:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	44.70
04/01/22 18:18	04/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.05
05/01/22 18:18	05/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.42
06/01/22 18:18	06/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.86
07/01/22 18:18	07/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.61
08/01/22 18:18	08/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.76
09/01/22 18:18	09/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.16
10/01/22 18:18	10/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.73
11/01/22 18:18	11/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.67
12/01/22 18:18	12/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.40
13/01/22 18:18	13/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.21
14/01/22 18:18	14/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.17
15/01/22 18:18	15/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.30
16/01/22 18:18	16/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.97
17/01/22 18:18	17/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.01
18/01/22 18:18	18/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.10
19/01/22 18:18	19/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.87
20/01/22 18:18	20/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.43

21/01/22 18:18	21/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.65
22/01/22 18:18	22/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.37
23/01/22 18:18	23/01/22 18:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.09
29/01/22 18:12	29/01/22 18:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.73
30/01/22 18:12	30/01/22 18:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.58
31/01/22 18:12	31/01/22 18:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.29

Recovery Boiler A

Monitoring Location No: 2
Monitoring Type: Continuous
Sample Type: Air
Description: Discharge duct downstream of Recovery Boiler A prior to junction with Stack 1

Nitrogen Oxides (as NO ₂)					
Period: 60 Minutes		Limit: 250.00 mg/Nm ³			
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
06/01/22 05:00	06/01/22 06:00	Instrument Error	Normal (Steady State)	Faulty Unit	498.55
06/01/22 20:00	06/01/22 21:00	Instrument Error	Normal (Steady State)	Faulty Unit	653.86
30/01/22 20:00	30/01/22 21:00	Instrument Error	Normal (Steady State)	Instrument Error	453.06

Authorised By:

Uday Bhagwat
Pulp Mill Manager

Johan Stoltz
General Manager


