

Visy Pulp and Paper Tumut GEMS - Correction Log

Reporting Period: 1/04/2021 - 1/05/2021

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity						Period: 6 Minutes	Limit: 20.00 %
Start Time	End Time	Cause	Operational State	Explanation	Max Reading		
01/04/21 14:30	01/04/21 14:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		23.89	
02/04/21 14:30	02/04/21 14:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		24.91	
03/04/21 14:30	03/04/21 14:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		21.26	
09/04/21 13:24	09/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		20.45	
10/04/21 13:24	10/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		20.52	
11/04/21 13:24	11/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		21.07	
12/04/21 13:24	12/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		21.94	
13/04/21 13:24	13/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		21.18	
14/04/21 13:24	14/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		21.59	
15/04/21 13:24	15/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		22.71	
16/04/21 13:24	16/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		22.86	
17/04/21 13:24	17/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		22.27	
18/04/21 13:24	18/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		22.48	
19/04/21 13:24	19/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		23.27	
20/04/21 13:24	20/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		23.21	
21/04/21 13:24	21/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		23.96	
22/04/21 13:24	22/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		24.76	
23/04/21 13:24	23/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		24.79	
24/04/21 13:24	24/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		24.89	
25/04/21 13:24	25/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration		24.94	

26/04/21 13:24	26/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.03
27/04/21 13:24	27/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.92
28/04/21 13:24	28/04/21 13:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.26

Power Boiler

Monitoring Location No: 3
Monitoring Type: Continuous
Sample Type: Air
Description: Discharge duct downstream of Power Boiler prior to junction with Stack 1

Carbon Monoxide (CO)		Period: 60 Minutes	Limit: 140.00 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
04/04/21 05:00	04/04/21 06:00	Instrument Error	Normal (Steady State)	Auto Calibration	150.59
22/04/21 09:00	22/04/21 11:00	Instrument Error	Normal (Steady State)	CO Probe calibrated by maintenance	164.72
23/04/21 17:00	23/04/21 18:00	Instrument Error	Normal (Steady State)	Calibration by maintenance	140.60

Authorised By:

Uday Bhagwat
Pulp Mill Manager

Johan Stoltz
General Manager


