

# Visy Pulp and Paper Tumut CEMS - Correction Log

Reporting Period:

1/05/2021 - 1/06/2021

Environment Protection Licence No:

10232

**Main Stack 1**

Monitoring Location No: 1  
 Monitoring Type: Continuous  
 Sample Type: Air  
 Description: Exit point from Stack 1 to atmosphere


Nitrogen Oxides (as NO2)		Period: 60 Minutes	Limit: 400.00 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
03/05/21 21:00	04/05/21 05:00	Instrument Error	Normal (Steady State)	Procal system refreshed and back online	457.52
04/05/21 05:00	04/05/21 07:00	Instrument Error	Normal (Steady State)	Procal system refreshed and back online	441.96

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
03/05/21 13:18	03/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.17
04/05/21 13:18	04/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.82
05/05/21 13:18	05/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.77
06/05/21 13:18	06/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.19
07/05/21 13:18	07/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.57
08/05/21 13:18	08/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.30
09/05/21 13:18	09/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.23
10/05/21 13:18	10/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.71
11/05/21 13:18	11/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.07
12/05/21 13:18	12/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.48
13/05/21 13:18	13/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.73
14/05/21 13:18	14/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.50
15/05/21 13:18	15/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.04
16/05/21 13:18	16/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.72

17/05/21 13:18	17/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.57
18/05/21 13:18	18/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.53
19/05/21 13:18	19/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.46
20/05/21 13:18	20/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.72
21/05/21 13:18	21/05/21 13:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.39
27/05/21 13:12	27/05/21 13:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.28
28/05/21 13:12	28/05/21 13:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.86
29/05/21 13:12	29/05/21 13:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.89
30/05/21 13:12	30/05/21 13:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.09
31/05/21 13:12	31/05/21 13:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.59

Authorised By:

**Uday Bhagwat**  
Pulp Mill Manager



**Johan Stoltz**  
General Manager

