



Visy Pulp and Paper Tumut CEMS - Correction Log

6/04/2022

Reporting Period: 1/03/2022 - 1/04/2022

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity						Period: 6 Minutes	Limit: 20.00 %
Start Time	End Time	Cause	Operational State	Explanation	Max Reading		
01/03/22 18:06	01/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.75		
02/03/22 18:06	02/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.03		
03/03/22 18:06	03/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.12		
04/03/22 18:06	04/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.38		
05/03/22 18:06	05/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.39		
06/03/22 18:06	06/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.83		
07/03/22 18:06	07/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.64		
08/03/22 18:06	08/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.01		
09/03/22 18:06	09/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.58		
10/03/22 18:06	10/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.04		
11/03/22 18:06	11/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.51		
12/03/22 18:06	12/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.71		
13/03/22 18:06	13/03/22 18:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.54		
18/03/22 18:00	18/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.59		
19/03/22 18:00	19/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.78		
20/03/22 18:00	20/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.75		
21/03/22 18:00	21/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.04		
22/03/22 18:00	22/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.70		

23/03/22 18:00	23/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.76
24/03/22 18:00	24/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.89
25/03/22 18:00	25/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.69
26/03/22 18:00	26/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.29
27/03/22 18:00	27/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.12
28/03/22 18:00	28/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.33
29/03/22 18:00	29/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.96
30/03/22 18:00	30/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.97
31/03/22 18:00	31/03/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.79

Power Boiler

Monitoring Location No: 3
Monitoring Type: Continuous
Sample Type: Air
Description: Discharge duct downstream of Power Boiler prior to junction with Stack 1

Carbon Monoxide (CO)		Period: 60 Minutes	Limit: 140.00 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
03/03/22 15:00	03/03/22 16:00	Instrument Error	Normal (Steady State)	Group Instruments calibration	151.87

Lime Kiln A


Monitoring Location No: 4
Monitoring Type: Continuous
Sample Type: Air
Description: Discharge duct downstream of Lime Kiln A prior to junction with Stack 1

Nitrogen Oxides (as NO2)		Period: 60 Minutes	Limit: 400.00 mg/Nm3		

Start Time	End Time	Cause	Operational State	Explanation	Max Reading
02/03/22 23:00	03/03/22 00:00	Instrument Error	Normal (Steady State)	Group Instruments working on system - Instrument went on fault	407.97
03/03/22 06:00	03/03/22 07:00	Instrument Error	Normal (Steady State)	Group Instruments working on system - Instrument went on fault	405.08

Authorised By:

Uday Bhagwat
Pulp Mill Manager


11-4-22

Johan Stoltz
General Manager

